

product	volume	kit size	Measuring Range	Method
<b>ASO</b> Anti streptolysin(O)	58 ml	1 x 8 mL Latex 2 x 25 mL Buffer	0 - 400 IU/mL	IMMUNOTURBIDIMETRY
<b>APO A1</b>	59ml	1 x 9 mL Antiserum 2 x 25 mL Buffer	0-300 mg/dL	IMMUNOTURBIDIMETRY
<b>APO A2</b>	58ml	1 x 8 mL Antiserum 2 x 25 mL Buffer	0-100 mg/dL	IMMUNOTURBIDIMETRY
<b>APO B</b>	55ml	1 x 5 mL Antiserum 2 x 25 mL Buffer	0-330 mg/dL	IMMUNOTURBIDIMETRY
<b>Complement C3</b>	58ml	1 x 8 mL Antiserum 2 x 25 mL Buffer	0-400 mg/dL	IMMUNOTURBIDIMETRY
<b>Complement C4</b>	58ml	1 x 8 mL Antiserum 2 x 25 mL Buffer	0-80 mg/dL	IMMUNOTURBIDIMETRY
<b>CRP US</b>	62ml	1 x 12 mL Latex 2 x 25 mL Buffer	0-14 mg/dL	IMMUNOTURBIDIMETRY
<b>CRP</b>	137ml	1 x 12 mL Antiserum 5 x 25 mL Buffer	0-22 mg/dL	IMMUNOTURBIDIMETRY
<b>Hb-A1c</b> Hemoglobin A1c	60ml	1 x 10 mL Hemoglobin A1c Reagent 2 2 x 25 mL Hemoglobin A1c Reagent 1	0-15 %	IMMUNOTURBIDIMETRY
<b>IgA</b>	58ml	1 x 8 mL Antiserum 2 x 25 mL Buffer	0-600 mg/dL	IMMUNOTURBIDIMETRY
<b>IgG</b>	60ml	1 x 10 mL Antiserum 2 x 25 mL Buffer	0-2700 mg/dL	IMMUNOTURBIDIMETRY
<b>IgM</b>	56ml	1 x 6 mL Antiserum 2 x 25 mL Buffer	0-500 mg/dL	IMMUNOTURBIDIMETRY
<b>Lipoprotein (a)</b>	59ml	1 x 9 mL Latex 2 x 25 mL Buffer	0-80 mg/dL	IMMUNOTURBIDIMETRY
<b>Microalbumin</b>	58ml	1 x 8 mL Antiserum 2 x 25 mL Buffer	0-400 mg/L	IMMUNOTURBIDIMETRY
<b>RF (3rd GEN)</b> Rheumatoid Factor	150ml	1 x 25 mL Reagent 5 x 25 mL Buffer	0-500 IU/mL	IMMUNOTURBIDIMETRY
<b>Transferrin</b>	55ml	1 x 5 mL Antiserum 2 x 25 mL Buffer	0-500 mg/dL	IMMUNOTURBIDIMETRY



## Your Reference in Turbidimetry

- Extensive list of application notes for currently available instruments
- CE-marked products
- ISO 9001:2000 and ISO 13485:2003 certified
- Anvisa Certified N° 349556/10-6
- Application approval on customer's and/or in-house instruments

product	vol	product	vol		
Hb A1C	Hemoglobin A1c Control High	0.5 ml	Transferrin	Transferrin Standard Set	5 x 1 ml
	Hemoglobin A1c Control Low	0.5 ml	APO	Apo A1/B Control	1 ml
	Hemoglobin A1c Standard Set	4 x 0.5 ml		Apo A1/B Standard High	1 ml
ASO	ASO Control	1 ml	Protein	Protein standard Set	5 x 1 ml
	ASO Standard Set	4 x 1 ml		Protein standard High	1 ml
	ASO Standard super High	1 ml	Immunology	Immunology Control High	1 ml
	ASO Standard High	1 ml		Immunology Control low	1 ml
CRP	CRP Control High	1 ml	Micro-Albumin	Microalbumin Control High	1 ml
	CRP Control low	1 ml		Microalbumin Control Low	1 ml
	CRP Standard High us	1 ml		Microalbumin Standard	1 ml
	CRP Standard High	1 ml		Microalbumin standard Set	5 x 1 ml
	CRP Standard Super High	1 ml	Fibrinogen	Fibrinogen Control	1 ml
	CRP Standard Set	5 x 1 ml		Fibrinogen Standard	0.5 ml
	RF Control	1 ml	LP(a)	LP (a)Control High	1 ml
RF	RF Standard Set	5 x 1 ml		LP (a)Control Low	1 ml
	RF Standard Super High	1 ml		LP (a) standard Set	4 x 1 ml
	RF Standard High	1 ml		LP (a) Standard High	1 ml
Ferritin	Ferritin Control High	1 ml			
	Ferritin Control Low	1 ml			
	Ferritin Standard High	5 x 1 ml			
	Ferritin Standard Set 2nd Generation	6 x 1 ml			

## Wako

product	volume	kit size	Measuring Range	Method
Auto kit CH50(wako)	60ml	2x20 ml liposome 1xfor 20ml substrate 1x20ml diluent	10-60 u/ml	Enzymatic
CH50 calibrator	5conc x for 0.5ml	CH50 calibrator	—	—
complement control	2conc x 1 bottlesx for 0.5ml	complement control	—	—
LDL-C (wako)	60ml	R1 : 1 x 45 ml R2 : 1 x 15 ml	1-400 mg/dL	Enzymatic Selective Protein method
HDL-C (wako)	60ml	R1 : 1 x 45 ml R2 : 1 x 15 ml	1-180 mg/dL	Immunoinhibition method
HDL-C/LDL-C calibrator	1ml	HDL-C/LDL-C calibrator	—	—



### Autokit CH50

Special characteristics:

Stable reagent

Homogeneous reagent free from precipitation

Good correlation with Mayer's method

Applicable to automated analyzers